

Steel Metric Gauge Block Sets Grade K with ISO JCSS Certificate



Calibration Certificate

Specifications

Accuracy | EN ISO 3650




Steel 112-block set



Steel 56-block set



Steel 34-block set



JCSS Calibration Certificate

These Gauge Block Sets are delivered with a JCSS Certificate of Calibration (Japan Calibration Service System).

A JCSS Certificate of Calibration is comparable with e.g. DAKKS COFRAC, RVA or UKAS Certificate of Calibration.

This calibration result is accepted internationally.

Grade K with ISO JCSS Certificate

Metric

Blocks per Set	No.	Grade	Block allocation					
			Size	Step	Quantity			
122	516-596-60	K	1,0005	-	1			
			1,001 - 1,009	0,001	9			
			1,01 - 1,49	0,01	49			
			1,6 - 1,9	0,1	4			
			0,5 - 24,5	0,5	49			
			30 - 100	10	8			
			25, 75	-	2			
112	516-937-60	K	1,0005	-	1			
			1,001 - 1,009	0,001	9			
			1,01 - 1,49	0,01	49			
			0,5 - 24,5	0,5	49			
			25 - 100	25	4			
103	516-941-60	K	1,005	-	1			
			1,01 - 1,49	0,01	49			
			0,5 - 24,5	0,5	49			
			25 - 100	25	4			
			87	516-945-60	K	1,001 - 1,009	0,001	9
1,01 - 1,49	0,01	49						
0,5 - 9,5	0,5	19						
10 - 100	10	10						
76	516-949-60	K				1,005	-	1
			1,01 - 1,49	0,01	49			
			0,5 - 9,5	0,5	19			
			10 - 40	10	4			
			50 - 100	25	3			
56	516-953-60	K	0,5	-	1			
			1,001 - 1,009	0,001	9			
			1,01 - 1,09	0,01	9			
			1,1 - 1,9	0,1	9			
			1 - 24	1	24			
			25 - 100	25	4			
			47	516-957-60	K	1,005	-	1
1,01 - 1,09	0,01	9						
1,1 - 1,9	0,1	9						
1 - 24	1	24						
25 - 100	25	4						
47	516-961-60	K	1,005	-	1			
			1,01 - 1,19	0,01	19			
			1,2 - 1,9	0,1	8			
			1 - 9	1	9			
			10 - 100	10	10			
46	516-994-60	K	1,001 - 1,009	0,001	9			
			1,01 - 1,09	0,01	9			
			1,1 - 1,9	0,1	9			
			1 - 9	1	9			
			10 - 100	10	10			
			34	516-128-60	K	1,0005	-	1
						1,001 - 1,009	0,001	9
1,01 - 1,09	0,01	9						
1,1 - 1,9	0,1	9						
1 - 5	1	5						
10	-	1						
32	516-965-60	K				1,005	-	1
			1,01 - 1,09	0,01	9			
			1,1 - 1,9	0,1	9			
			1 - 9	1	9			
			10 - 30	10	3			
			60	-	1			

Steel Metric Gauge Block Sets Grade 0 with ISO JCSS Certificate

Grade 0 with ISO JCSS Certificate



Calibration Certificate

Metric

Blocks per Set	No.	Grade	Block allocation		
			Size	Step	Quantity
122	516-597-60	0	1,0005	-	1
			1,001 - 1,009	0,001	9
			1,01 - 1,49	0,01	49
			1,6 - 1,9	0,1	4
			0,5 - 24,5	0,5	49
			30 - 100	10	8
			25, 75	-	2
112	516-938-60	0	1,0005	-	1
			1,001 - 1,009	0,001	9
			1,01 - 1,49	0,01	49
			0,5 - 24,5	0,5	49
			25 - 100	25	4
103	516-942-60	0	1,005	-	1
			1,01 - 1,49	0,01	49
			0,5 - 24,5	0,5	49
			25 - 100	25	4
88	516-970-60	0	1,0005	-	1
			1,001 - 1,009	0,001	9
			1,01 - 1,49	0,01	49
			0,5 - 9,5	0,5	19
			10 - 100	10	10
87	516-946-60	0	1,001 - 1,009	0,001	9
			1,01 - 1,49	0,01	49
			0,5 - 9,5	0,5	19
			10 - 100	10	10
76	516-950-60	0	1,005	-	1
			1,01 - 1,49	0,01	49
			0,5 - 9,5	0,5	19
			10 - 40	10	4
			50 - 100	25	3
56	516-954-60	0	0,5	-	1
			1,001 - 1,009	0,001	9
			1,01 - 1,09	0,01	9
			1,1 - 1,9	0,1	9
			1 - 24	1	24
			25 - 100	25	4
47	516-958-60	0	1,005	-	1
			1,01 - 1,09	0,01	9
			1,1 - 1,9	0,1	9
			1 - 24	1	24
			25 - 100	25	4
47	516-962-60	0	1,005	-	1
			1,01 - 1,19	0,01	19
			1,2 - 1,9	0,1	8
			1 - 9	1	9
			10 - 100	10	10
46	516-995-60	0	1,001 - 1,009	0,001	9
			1,01 - 1,09	0,01	9
			1,1 - 1,9	0,1	9
			1 - 9	1	9
			10 - 100	10	10
34	516-129-60	0	1,0005	-	1
			1,001 - 1,009	0,001	9
			1,01 - 1,09	0,01	9
			1,1 - 1,9	0,1	9
			1 - 5	1	5
			10	-	1
32	516-966-60	0	1,005	-	1
			1,01 - 1,09	0,01	9
			1,1 - 1,9	0,1	9
			1 - 9	1	9
			10 - 30	10	3
			60	-	1

Specifications

Accuracy EN ISO 3650




Steel 122-block set



Steel 88-block set



Steel 87-block set



JCSS Calibration Certificate

These Gauge Block Sets are delivered with a JCSS Certificate of Calibration (Japan Calibration Service System).

A JCSS Certificate of Calibration is comparable with e.g. DAKS COFRAC, RVA or UKAS Certificate of Calibration.

This calibration result is accepted internationally.

Steel Metric Gauge Block Sets Grade 1 with ISO JCSS Certificate

Specifications

Accuracy | EN ISO 3650




Steel 103-block set



Steel 76-block set



Steel 46-block set



JCSS Calibration Certificate
These Gauge Block Sets are delivered with a JCSS Certificate of Calibration (Japan Calibration Service System).
A JCSS Certificate of Calibration is comparable with e.g. DAKKS COFRAC, RVA or UKAS Certificate of Calibration.
This calibration result is accepted internationally.

Grade 1 with ISO JCSS Certificate



Calibration Certificate

Metric

Blocks per Set	No.	Grade	Block allocation		
			Size	Step	Quantity
122	516-598-60	1	1,0005	-	1
			1,001 - 1,009	0,001	9
			1,01 - 1,49	0,01	49
			1,6 - 1,9	0,1	4
			0,5 - 24,5	0,5	49
			30 - 100	10	8
			25, 75	-	2
112	516-939-60	1	1,0005	-	1
			1,001 - 1,009	0,001	9
			1,01 - 1,49	0,01	49
			0,5 - 24,5	0,5	49
			25 - 100	25	4
103	516-943-60	1	1,005	-	1
			1,01 - 1,49	0,01	49
			0,5 - 24,5	0,5	49
			25 - 100	25	4
88	516-971-60	1	1,0005	-	1
			1,001 - 1,009	0,001	9
			1,01 - 1,49	0,01	49
			0,5 - 9,5	0,5	19
			10 - 100	10	10
87	516-947-60	1	1,001 - 1,009	0,001	9
			1,01 - 1,49	0,01	49
			0,5 - 9,5	0,5	19
			10 - 100	10	10
76	516-951-60	1	1,005	-	1
			1,01 - 1,49	0,01	49
			0,5 - 9,5	0,5	19
			10 - 40	10	4
			50 - 100	25	3
56	516-955-60	1	0,5	-	1
			1,001 - 1,009	0,001	9
			1,01 - 1,09	0,01	9
			1,1 - 1,9	0,1	9
			1 - 24	1	24
			25 - 100	25	4
47	516-959-60	1	1,005	-	1
			1,01 - 1,09	0,01	9
			1,1 - 1,9	0,1	9
			1 - 24	1	24
			25 - 100	25	4
47	516-963-60	1	1,005	-	1
			1,01 - 1,19	0,01	19
			1,2 - 1,9	0,1	8
			1 - 9	1	9
			10 - 100	10	10
46	516-996-60	1	1,001 - 1,009	0,001	9
			1,01 - 1,09	0,01	9
			1,1 - 1,9	0,1	9
			1 - 9	1	9
			10 - 100	10	10
34	516-130-60	1	1,0005	-	1
			1,001 - 1,009	0,001	9
			1,01 - 1,09	0,01	9
			1,1 - 1,9	0,1	9
			1 - 5	1	5
			10	-	1
32	516-967-60	1	1,005	-	1
			1,01 - 1,09	0,01	9
			1,1 - 1,9	0,1	9
			1 - 9	1	9
			10 - 30	10	3
			60	-	1

Steel Metric Gauge Block Sets Grade 2 with ISO JCSS Certificate

Grade 2 with ISO JCSS Certificate



Calibration Certificate

Metric

Blocks per Set	No.	Grade	Block allocation		
			Size	Step	Quantity
122	516-599-60	2	1,0005	-	1
			1,001 - 1,009	0,001	9
			1,01 - 1,49	0,01	49
			1,6 - 1,9	0,1	4
			0,5 - 24,5	0,5	49
			30 - 100	10	8
			25, 75	-	2
112	516-940-60	2	1,0005	-	1
			1,001 - 1,009	0,001	9
			1,01 - 1,49	0,01	49
			0,5 - 24,5	0,5	49
			25 - 100	25	4
103	516-944-60	2	1,005	-	1
			1,01 - 1,49	0,01	49
			0,5 - 24,5	0,5	49
			25 - 100	25	4
88	516-972-60	2	1,0005	-	1
			1,001 - 1,009	0,001	9
			1,01 - 1,49	0,01	49
			0,5 - 9,5	0,5	19
			10 - 100	10	10
87	516-948-60	2	1,001 - 1,009	0,001	9
			1,01 - 1,49	0,01	49
			0,5 - 9,5	0,5	19
			10 - 100	10	10
76	516-952-60	2	1,005	-	1
			1,01 - 1,49	0,01	49
			0,5 - 9,5	0,5	19
			10 - 40	10	4
			50 - 100	25	3
56	516-956-60	2	0,5	-	1
			1,001 - 1,009	0,001	9
			1,01 - 1,09	0,01	9
			1,1 - 1,9	0,1	9
			1 - 24	1	24
			25 - 100	25	4
47	516-960-60	2	1,005	-	1
			1,01 - 1,09	0,01	9
			1,1 - 1,9	0,1	9
			1 - 24	1	24
			25 - 100	25	4
47	516-964-60	2	1,005	-	1
			1,01 - 1,19	0,01	19
			1,2 - 1,9	0,1	8
			1 - 9	1	9
			10 - 100	10	10
46	516-997-60	2	1,001 - 1,009	0,001	9
			1,01 - 1,09	0,01	9
			1,1 - 1,9	0,1	9
			1 - 9	1	9
			10 - 100	10	10
34	516-131-60	2	1,0005	-	1
			1,001 - 1,009	0,001	9
			1,01 - 1,09	0,01	9
			1,1 - 1,9	0,1	9
			1 - 5	1	5
			10	-	1
32	516-968-60	2	1,005	-	1
			1,01 - 1,09	0,01	9
			1,1 - 1,9	0,1	9
			1 - 9	1	9
			10 - 30	10	3
			60	-	1

Specifications

Accuracy EN ISO 3650




Steel 112-block set



Steel 47-block set



Steel 32-block set



JCSS Calibration Certificate

These Gauge Block Sets are delivered with a JCSS Certificate of Calibration (Japan Calibration Service System).

A JCSS Certificate of Calibration is comparable with e.g. DAKKS COFRAC, RVA or UKAS Certificate of Calibration.

This calibration result is accepted internationally.