

Ceramic Gauge Block Sets

Series 516 - 1 Base

Ordering example: To order an ISO Standard 112-block grade 1 set with JCSS calibration certificate choose 516-339-60.



Calibration
Certificate



Inspection
Certificate

Metric

Blocks per Set	No.	Standard / grade available and Suffix No. *		Block allocation	
		Size		Step	Quantity
112	516-337	1,0005	-	1	1
	516-338	1,001 - 1,009	0,001	9	9
	516-339	1,01 - 1,49	0,01	49	49
	516-340	0,5 - 24,5	0,5	49	49
103	516-341	25 - 100	25	4	4
	516-342	1,005	-	1	1
	516-343	1,01 - 1,49	0,01	49	49
	516-344	0,5 - 24,5	0,5	49	49
88	516-370	25 - 100	25	4	4
	516-371	1,0005	-	1	1
	516-372	1,001 - 1,009	0,001	9	9
	516-373	1,01 - 1,49	0,01	49	49
87	516-374	0,5 - 9,5	0,5	19	19
	516-375	10 - 100	10	10	10
	516-376	1,001 - 1,009	0,001	9	9
	516-377	1,01 - 1,49	0,01	49	49
76	516-378	0,5 - 9,5	0,5	19	19
	516-379	10 - 100	10	10	10
	516-380	1,005	-	1	1
	516-381	1,01 - 1,49	0,01	49	49
56	516-382	0,5 - 9,5	0,5	19	19
	516-383	10 - 40	10	4	4
	516-384	50 - 100	25	3	3
	516-385	0,5	-	1	1
47	516-386	1,001 - 1,009	0,001	9	9
	516-387	1,01 - 1,09	0,01	9	9
	516-388	1,1 - 1,9	0,1	9	9
	516-389	1 - 24	1	24	24
47	516-390	25 - 100	25	4	4
	516-391	1,005	-	1	1
	516-392	1,01 - 1,19	0,01	19	19
	516-393	1,2 - 1,9	0,1	8	8
46	516-394	1 - 9	1	9	9
	516-395	10 - 100	10	10	10
	516-396	1,001 - 1,009	0,001	9	9
	516-397	1,01 - 1,09	0,01	9	9
34	516-398	1,1 - 1,9	0,1	9	9
	516-399	1 - 9	1	9	9
	516-400	10 - 100	10	10	10
	516-401	1 - 5	1	5	5
32	516-402	10	-	1	1
	516-403	1,0005	-	1	1
	516-404	1,001 - 1,009	0,001	9	9
	516-405	1,01 - 1,09	0,01	9	9
8	516-406	1,1 - 1,9	0,1	9	9
	516-407	1 - 9	1	9	9
	516-408	10 - 30	10	3	3
	516-409	60	-	1	1
8	516-731	125 - 175	25	3	3
	516-732	200 - 250	50	2	2
	516-733	300 - 500	100	3	3
	516-734				

Specifications

Accuracy EN ISO 3650

Suffix No. (-X) for Selecting Certificate Provided

ISO/DIN/JIS		
Suffix No.	Inspection Certificate	Calibration Certificate JCSS
1	○	—
6	○	○

Suffix No. 1: Not available for Grade K sets.



CERA 112-block set



CERA 56-block set



CERA 32-block set